
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 003564
8:07 02/05/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:42
PRESSURIZING 2:46
STERILIZING 3:52
136.4 C 222kPa 3:52
136.4 C 223kPa 4:22
136.1 C 222kPa 4:52
136.2 C 222kPa 5:22
136.4 C 222kPa 5:52
136.4 C 223kPa 6:22
136.4 C 222kPa 6:52
136.3 C 222kPa 7:22
136.3 C 222kPa 7:22
VENTING 7:22
AIR DRYING 8:14
CYCLE COMPLETE 9:40

Digital Signature #
017F0E401CA4AFDE
