

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3283

Name: RITTER2  
Sterilizer ID: V2313228  
01/10/2024 12:51 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 153.9 F 0.0  
2:00 154.8 F 1.7  
4:00 172.5 F 5.1  
6:00 195.1 F 9.4  
8:00 212.9 F 9.0  
10:00 220.5 F 9.1  
12:00 232.6 F 10.6  
14:00 236.9 F 11.7  
16:00 237.7 F 10.3  
18:00 250.4 F 16.6  
20:00 261.2 F 23.0  
22:00 271.6 F 29.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 30.1  
1:00 272.6 F 30.7  
2:00 271.5 F 30.0  
3:00 272.5 F 30.9  
4:00 272.2 F 30.5

Min 271.2 F 30.0  
Max 272.8 F 30.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.0 F 0.0  
5:00 234.6 F 0.0  
10:00 224.2 F 0.0  
15:00 183.4 F 0.0  
20:00 178.8 F 0.0  
25:00 177.5 F 0.0  
30:00 176.8 F 0.0  
Drying Complete

FILLING: 1:06  
HEATING: 23:15  
STERILIZING: 4:00  
VENTING: 3:44  
DRYING: 30:00  
TOTAL CYCLE: 01:02:06  
CYCLE COMPLETE

01/10/2024 01:53 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_