
Midmark M11 - v1.0.0
Total Cycles: 1740

Name: RITTER 1
Sterilizer ID: V2178198
03/24/2022 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 174.1 F 0.0
2:00 167.1 F 2.7
4:00 188.2 F 6.8
6:00 208.3 F 12.1
8:00 226.7 F 6.9
10:00 238.4 F 12.2
12:00 236.5 F 9.4
14:00 237.3 F 9.0
16:00 248.6 F 14.5
18:00 261.0 F 21.5
20:00 272.3 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.7
1:00 271.7 F 28.9
2:00 271.8 F 29.0
3:00 271.9 F 29.0
4:00 271.7 F 28.9

Min 271.4 F 28.6
Max 272.6 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.6 F 0.0
5:00 231.0 F 0.0
10:00 230.9 F 0.0
15:00 228.2 F 0.0
20:00 230.4 F 0.0
25:00 235.4 F 0.0
30:00 220.1 F 0.0
Drying Complete

FILLING: 1:20
HEATING: 20:22
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 00:58:52
CYCLE COMPLETE

03/24/2022 11:31 AM
ACCEPTED: X
REJECTED:
Operator: U02