

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1122

Name: RITTER2  
Sterilizer ID: V2313228  
05/26/2022 11:14 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 146.7 F 0.0  
2:00 153.2 F 1.7  
4:00 177.0 F 5.2  
6:00 207.7 F 10.2  
8:00 216.0 F 5.6  
10:00 223.7 F 10.0  
12:00 232.3 F 8.5  
14:00 239.4 F 12.5  
16:00 239.9 F 11.2  
18:00 252.9 F 17.8  
20:00 264.4 F 24.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.5  
1:00 271.8 F 30.2  
2:00 272.3 F 29.8  
3:00 271.4 F 29.7  
4:00 272.6 F 30.2  
  
Min 271.2 F 29.3  
Max 272.7 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 185.0 F 0.0  
5:00 202.1 F 0.0  
10:00 212.6 F 0.0  
15:00 217.7 F 0.0  
20:00 217.7 F 0.0  
25:00 227.2 F 0.0  
30:00 215.9 F 0.0  
Drying Complete

FILLING: 1:25  
HEATING: 21:48  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:00:21  
CYCLE COMPLETE

05/26/2022 12:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_