
Midmark M11 - v1.0.0
Total Cycles: 5397

Name: RITTER 1
Sterilizer ID: V2178198
03/07/2025 10:09 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.4 F 0.0
2:00 89.1 F 0.3
4:00 110.8 F 1.7
6:00 137.5 F 4.1
8:00 153.0 F 7.0
10:00 180.8 F 10.6
12:00 204.2 F 4.2
14:00 214.4 F 5.0
16:00 224.2 F 8.4
18:00 233.8 F 12.2
20:00 230.7 F 7.4
22:00 242.1 F 12.4
24:00 238.3 F 9.9
26:00 251.0 F 16.0
28:00 261.9 F 22.2
30:00 272.1 F 29.0

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.5
1:00 272.0 F 29.0
2:00 272.5 F 28.9
3:00 271.5 F 28.6
4:00 272.7 F 29.3

Min 271.4 F 28.3
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.4 F 0.0
5:00 188.0 F 0.0
10:00 194.6 F 0.0
15:00 198.7 F 0.0
20:00 197.1 F 0.0
25:00 202.8 F 0.0
30:00 205.6 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 30:24
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:08:40
CYCLE COMPLETE

03/07/2025 11:17 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____