
Midmark M11 - v1.0.0
Total Cycles: 1443

Name: RITTER 1
Sterilizer ID: V2178198
12/09/2021 01:37 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.1 F 0.0
2:00 198.5 F 3.6
4:00 208.1 F 9.2
6:00 224.8 F 6.9
8:00 235.0 F 10.7
10:00 234.0 F 8.1
12:00 240.4 F 10.1
14:00 248.2 F 14.3
16:00 261.3 F 21.5
18:00 272.7 F 28.9

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.2
1:00 272.7 F 28.8
2:00 271.9 F 28.5
3:00 272.0 F 28.4
4:00 272.6 F 29.0

Min 271.5 F 28.0
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 196.1 F 0.0
10:00 202.7 F 0.0
15:00 205.0 F 0.0
20:00 205.7 F 0.0
25:00 205.2 F 0.0
30:00 205.2 F 0.0
Drying Complete

FILLING: 3:06
HEATING: 18:12
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 00:58:24
CYCLE COMPLETE

12/09/2021 02:46 PM
ACCEPTED: X
REJECTED:
Operator: U02