

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3049

Name: RITTER2  
Sterilizer ID: V2313228  
10/24/2023 11:43 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 145.4 F 0.0  
2:00 144.9 F 2.0  
4:00 162.3 F 5.4  
6:00 187.1 F 10.0  
8:00 222.8 F 6.0  
10:00 229.8 F 9.9  
12:00 232.5 F 8.1  
14:00 241.0 F 12.3  
16:00 240.5 F 10.8  
18:00 253.5 F 17.4  
20:00 265.1 F 24.4  
22:00 271.7 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 29.3  
1:00 271.9 F 29.2  
2:00 272.3 F 29.4  
3:00 272.3 F 29.7  
4:00 272.2 F 29.9

Min 271.5 F 29.2  
Max 272.7 F 29.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.8 F 0.0  
5:00 246.6 F 0.0  
10:00 197.0 F 0.0  
15:00 194.7 F 0.0  
20:00 190.8 F 0.0  
25:00 188.8 F 0.0  
30:00 187.5 F 0.0  
Drying Complete

FILLING: 1:01  
HEATING: 22:29  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:00:46  
CYCLE COMPLETE

10/24/2023 12:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_