
Midmark M11 - v1.0.0
Total Cycles: 1572

Name: RITTER 1
Sterilizer ID: V2178198
01/31/2022 11:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 179.1 F 0.0
2:00 178.6 F 3.3
4:00 198.7 F 8.2
6:00 219.2 F 9.0
8:00 228.5 F 8.9
10:00 237.6 F 10.5
12:00 242.2 F 12.0
14:00 240.9 F 10.6
16:00 254.6 F 17.5
18:00 266.1 F 24.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.2
1:00 272.5 F 28.8
2:00 271.6 F 28.3
3:00 272.4 F 28.6
4:00 271.6 F 28.4

Min 271.4 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.2 F 0.0
5:00 236.1 F 0.0
10:00 227.9 F 0.0
15:00 229.9 F 0.0
20:00 228.2 F 0.0
25:00 230.0 F 0.0
30:00 230.2 F 0.0

Drying Complete

FILLING: 1:48
HEATING: 19:29
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:58:28
CYCLE COMPLETE

01/31/2022 12:46 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____