
Midmark M11 - v1.0.0
Total Cycles: 3306

Name: RITTER 1
Sterilizer ID: V2178198
05/30/2023 01:52 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 206.3 F 0.0
2:00 204.5 F 3.0
4:00 227.6 F 9.1
6:00 235.3 F 8.5
8:00 240.4 F 12.1
10:00 239.7 F 10.4
12:00 238.1 F 9.2
14:00 252.7 F 16.4
16:00 264.7 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.0
1:00 272.7 F 28.8
2:00 272.0 F 28.2
3:00 271.6 F 28.2
4:00 272.8 F 28.9

Min 271.3 F 27.8
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.5 F 0.0
5:00 231.4 F 0.0
10:00 229.7 F 0.0
15:00 229.6 F 0.0
20:00 229.4 F 0.0
25:00 222.5 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 2:51
HEATING: 17:56
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:04
CYCLE COMPLETE

05/30/2023 02:50 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____