
Midmark M11 - v1.0.0
Total Cycles: 5422

Name: RITTER 1
Sterilizer ID: V2178198
03/19/2025 09:43 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.6 F 0.0
2:00 174.6 F 2.1
4:00 174.9 F 5.8
6:00 199.3 F 9.8
8:00 218.6 F 5.2
10:00 224.6 F 7.5
12:00 234.9 F 12.0
14:00 234.0 F 8.8
16:00 241.2 F 10.7
18:00 245.1 F 13.0
20:00 257.0 F 19.2
22:00 267.3 F 25.6

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 272.0 F 29.3
2:00 271.8 F 28.7
3:00 271.4 F 28.9
4:00 271.9 F 29.1

Min 271.0 F 28.4
Max 272.2 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.9 F 0.0
5:00 206.7 F 0.0
10:00 214.8 F 0.0
15:00 220.7 F 0.0
20:00 230.0 F 0.0
25:00 219.6 F 0.0
30:00 226.1 F 0.0
Drying Complete

FILLING: 1:10
HEATING: 23:21
STERILIZING: 4:00
VENTING: 3:47
DRYING: 30:00
TOTAL CYCLE: 01:02:18
CYCLE COMPLETE

03/19/2025 10:45 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____