
Midmark M11 - v1.0.5
Total Cycles: 2914

Name: RITTER2
Sterilizer ID: V2313228
09/21/2023 09:56 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 139.6 F 0.0
2:00 136.9 F 1.0
4:00 154.1 F 3.8
6:00 179.3 F 8.3
8:00 203.8 F 13.5
10:00 220.3 F 6.5
12:00 238.0 F 12.1
14:00 237.3 F 9.7
16:00 237.7 F 9.0
18:00 249.3 F 15.1
20:00 261.3 F 21.8
22:00 272.0 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 272.3 F 29.1
2:00 271.6 F 29.1
3:00 271.6 F 28.5
4:00 272.3 F 29.1

Min 271.1 F 28.4
Max 272.4 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.1 F 0.0
5:00 244.6 F 0.0
10:00 218.2 F 0.0
15:00 209.3 F 0.0
20:00 177.2 F 0.0
25:00 195.7 F 0.0
30:00 223.0 F 0.0
Drying Complete

FILLING: 0:58
HEATING: 22:22
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

09/21/2023 10:56 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____