
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 001753
14:41 24/04/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:52
PRESSURIZING 1:58
STERILIZING 3:02
136.4 C 222kPa 3:02
136.2 C 223kPa 3:32
136.3 C 221kPa 4:02
136.3 C 225kPa 4:32
136.2 C 222kPa 5:02
136.4 C 225kPa 5:32
136.4 C 224kPa 6:02
136.4 C 225kPa 6:32
136.4 C 225kPa 6:32
VENTING 6:32
AIR DRYING 7:20
CYCLE COMPLETE 15:20

Digital Signature #
2D69C1D29CF409D2
