

-----  
Midmark M11 - v1.0.5  
Total Cycles: 442

Name: RITTER2  
Sterilizer ID: V2313228  
10/13/2021 06:23 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 74.6 F 0.0  
2:00 74.8 F 0.0  
4:00 80.4 F 0.0  
6:00 106.7 F 0.2  
8:00 140.8 F 2.0  
10:00 174.4 F 5.9  
12:00 196.0 F 10.1  
14:00 213.1 F 7.6  
16:00 222.7 F 8.2  
18:00 232.4 F 12.6  
20:00 233.9 F 9.4  
22:00 239.0 F 10.1  
24:00 245.1 F 13.5  
26:00 257.0 F 19.8  
28:00 267.6 F 26.3  
30:00 272.4 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.9 F 29.5  
1:00 272.2 F 29.4  
2:00 272.4 F 29.5  
3:00 272.6 F 29.7  
4:00 272.8 F 30.0

Min 271.8 F 29.4  
Max 273.1 F 30.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 187.3 F 0.0  
5:00 231.1 F 0.0  
10:00 199.0 F 0.0  
15:00 187.4 F 0.0  
20:00 182.2 F 0.0  
25:00 200.4 F 0.0  
30:00 190.6 F 0.0  
Drying Complete

FILLING: 0:59  
HEATING: 30:06  
STERILIZING: 4:00  
VENTING: 2:55  
DRYING: 30:00  
TOTAL CYCLE: 01:08:01  
CYCLE COMPLETE

10/13/2021 07:31 AM  
ACCEPTED: X  
REJECTED:  
Operator: U01