
Midmark M11 - v1.0.0
Total Cycles: 4082

Name: RITTER 1
Sterilizer ID: V2178198
12/05/2023 08:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 70.6 F 0.0
2:00 70.8 F 0.2
4:00 81.3 F 1.0
6:00 112.4 F 2.9
8:00 142.5 F 5.4
10:00 162.7 F 8.2
12:00 187.0 F 11.6
14:00 207.5 F 2.3
16:00 219.6 F 5.6
18:00 226.6 F 8.9
20:00 233.6 F 12.6
22:00 232.8 F 8.0
24:00 244.0 F 12.8
26:00 240.3 F 10.4
28:00 251.7 F 16.1
30:00 261.7 F 21.8
32:00 271.3 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 28.6
1:00 272.7 F 29.1
2:00 271.8 F 28.9
3:00 271.8 F 28.6
4:00 272.3 F 29.1

Min 271.3 F 28.5
Max 273.1 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 175.2 F 0.0
5:00 245.6 F 0.0
10:00 203.6 F 0.0
15:00 214.8 F 0.0
20:00 202.4 F 0.0
25:00 197.3 F 0.0
30:00 183.6 F 0.0

Drying Complete

FILLING: 0:51
HEATING: 33:13
STERILIZING: 4:00
VENTING: 3:38
DRYING: 30:00
TOTAL CYCLE: 01:11:42
CYCLE COMPLETE

12/05/2023 10:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____