
Midmark M11 - v1.0.0
Total Cycles: 1884

Name: RITTER 1
Sterilizer ID: V2178198
04/28/2022 09:51 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.4 F 0.0
2:00 191.3 F 3.2
4:00 188.9 F 7.3
6:00 211.6 F 12.3
8:00 220.6 F 6.8
10:00 234.2 F 12.1
12:00 235.4 F 9.2
14:00 238.5 F 9.2
16:00 248.0 F 14.2
18:00 260.3 F 21.1
20:00 271.5 F 28.3

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.5 F 28.9
2:00 271.6 F 28.5
3:00 272.5 F 28.9
4:00 271.5 F 28.3

Min 271.5 F 28.2
Max 272.7 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.4 F 0.0
5:00 209.1 F 0.0
10:00 224.2 F 0.0
15:00 231.7 F 0.0
20:00 230.5 F 0.0
25:00 226.9 F 0.0
30:00 231.3 F 0.0
Drying Complete

FILLING: 1:13
HEATING: 20:38
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:59
CYCLE COMPLETE

04/28/2022 10:49 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____