
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 008134
11:18 03/12/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:45
PRESSURIZING 3:45
STERILIZING 4:49
136.5 C 220kPa 4:49
136.3 C 224kPa 5:19
136.4 C 220kPa 5:49
136.4 C 221kPa 6:19
136.3 C 223kPa 6:49
136.4 C 224kPa 7:19
136.4 C 224kPa 7:49
136.4 C 223kPa 8:19
136.4 C 223kPa 8:19
VENTING 8:20
AIR DRYING 9:22
CYCLE COMPLETE 9:24

Digital Signature #
8178F8C6A52CB16C
