

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4178

Name: RITTER2  
Sterilizer ID: V2313228  
12/03/2024 03:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 100.8 F 0.0  
2:00 101.4 F 0.0  
4:00 116.4 F 0.2  
6:00 147.4 F 1.6  
8:00 174.9 F 4.5  
10:00 197.7 F 7.8  
12:00 204.7 F 11.9  
14:00 219.1 F 6.5  
16:00 230.6 F 11.2  
18:00 231.9 F 8.3  
20:00 241.3 F 11.9  
22:00 244.8 F 13.1  
24:00 256.3 F 19.3  
26:00 266.9 F 25.9  
28:00 273.2 F 30.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 29.6  
1:00 272.8 F 29.9  
2:00 271.8 F 30.0  
3:00 272.6 F 29.8  
4:00 272.0 F 30.1

Min 271.6 F 29.5  
Max 273.0 F 30.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 190.7 F 0.0  
5:00 213.7 F 0.0  
10:00 208.0 F 0.0  
15:00 195.5 F 0.0  
20:00 205.6 F 0.0  
25:00 192.7 F 0.0  
30:00 193.5 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 28:19  
STERILIZING: 4:00  
VENTING: 3:50  
DRYING: 30:00  
TOTAL CYCLE: 01:07:14  
CYCLE COMPLETE

12/03/2024 04:38 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_