
Midmark M11 - v1.0.5
Total Cycles: 819

Name: RITTER2
Sterilizer ID: V2313228
02/22/2022 10:38 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.1 F 0.0
2:00 183.8 F 3.1
4:00 200.6 F 9.6
6:00 218.1 F 5.0
8:00 240.5 F 12.7
10:00 244.3 F 13.2
12:00 246.3 F 13.6
14:00 262.8 F 22.8

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.6
1:00 272.7 F 29.4
2:00 272.6 F 28.9
3:00 272.0 F 28.5

Min 271.8 F 28.4
Max 272.9 F 29.5

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.7 F 0.0
Dry Cycle Aborted

FILLING: 1:53
HEATING: 15:40
STERILIZING: 3:00
VENTING: 2:23
DRYING: 1:26
TOTAL CYCLE: 00:24:22

02/22/2022 11:02 AM

ACCEPTED:

REJECTED:

Reason: -----
Operator: -----