
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 004403
8:57 21/09/2022

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:09
PRESSURIZING 2:11
STERILIZING 3:15
136.5 C 224kPa 3:15
136.2 C 222kPa 3:45
136.3 C 220kPa 4:15
136.4 C 224kPa 4:45
136.3 C 219kPa 5:15
136.2 C 222kPa 5:45
136.4 C 223kPa 6:15
136.4 C 223kPa 6:45
136.4 C 223kPa 6:45
VENTING 6:45
AIR DRYING 7:34
CYCLE COMPLETE 8:09

Digital Signature #
ECCCECBCA5C729F0
