
Midmark M11 - v1.0.0
Total Cycles: 1497

Name: RITTER 1
Sterilizer ID: V2178198
01/05/2022 01:02 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 162.4 F 0.0
2:00 167.2 F 2.3
4:00 178.5 F 6.1
6:00 199.5 F 10.8
8:00 214.7 F 3.9
10:00 226.4 F 9.0
12:00 237.6 F 10.9
14:00 239.5 F 11.2
16:00 237.0 F 9.3
18:00 250.6 F 15.8
20:00 262.3 F 22.5
22:00 272.7 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.5
1:00 272.4 F 29.3
2:00 271.6 F 28.6
3:00 271.9 F 29.1
4:00 272.4 F 29.1

Min 271.1 F 28.4
Max 272.5 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.7 F 0.0
5:00 213.2 F 0.0
10:00 217.3 F 0.0
15:00 222.8 F 0.0
20:00 228.5 F 0.0
25:00 231.1 F 0.0
30:00 230.1 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 22:23
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:34
CYCLE COMPLETE

01/05/2022 02:02 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____