

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.0uS / 0.0ppm  
CYCLE NUMBER 006336  
7:50 25/09/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:54  
PRESSURIZING 2:56  
STERILIZING 4:02  
136.0 C 220kPa 4:02  
136.4 C 221kPa 4:32  
136.2 C 222kPa 5:02  
136.0 C 219kPa 5:32  
136.3 C 222kPa 6:02  
136.2 C 221kPa 6:32  
136.2 C 221kPa 7:02  
136.2 C 221kPa 7:32  
136.2 C 221kPa 7:32  
VENTING 7:32  
AIR DRYING 8:23  
CYCLE COMPLETE 10:15

-----  
Digital Signature #  
0D03A53BF951E8D7  
-----