

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 000780  
8:37 29/09/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:28  
PRESSURIZING 2:35  
STERILIZING 3:39  
136.4 C 233kPa 3:39  
136.2 C 232kPa 4:09  
136.2 C 232kPa 4:39  
136.4 C 233kPa 5:09  
136.1 C 229kPa 5:39  
136.3 C 232kPa 6:09  
136.4 C 233kPa 6:39  
136.4 C 233kPa 7:09  
136.4 C 233kPa 7:09  
VENTING 7:09  
AIR DRYING 8:04  
CYCLE COMPLETE 8:06  
-----

Digital Signature #  
E13E03577EEE123B  
-----