
Midmark M11 - v1.0.0
Total Cycles: 1730

Name: RITTER 1
Sterilizer ID: V2178198
03/21/2022 11:59 AM

BEGIN UNWRAPPED CYCLE
Temp: 270 Degrees F
Time: 3 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	196.0 F	0.0
2:00	192.7 F	4.2
4:00	209.4 F	10.0
6:00	222.7 F	5.5
8:00	241.3 F	13.5
10:00	243.8 F	12.4
12:00	243.9 F	11.9
14:00	258.7 F	19.7

C103: HEATUP MODE
STOP PRESSED
CHAMBER TEMP 243F
CHAMBER PRESS 11.6PSI
** ITEMS NOT STERILE **

03/21/2022 12:16 PM