
Midmark M11 - v1.0.0
Total Cycles: 3861

Name: RITTER 1
Sterilizer ID: V2178198
10/04/2023 12:45 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 138.1 F 0.0
2:00 142.3 F 1.8
4:00 162.1 F 5.1
6:00 187.7 F 8.5
8:00 201.8 F 12.5
10:00 218.1 F 5.9
12:00 229.2 F 10.2
14:00 234.2 F 8.5
16:00 240.3 F 11.6
18:00 237.8 F 9.5
20:00 250.5 F 15.6
22:00 261.6 F 21.9
24:00 271.7 F 28.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.4
1:00 272.0 F 28.6
2:00 272.4 F 29.0
3:00 271.3 F 28.5
4:00 272.1 F 28.7

Min 271.3 F 28.3
Max 272.4 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 202.1 F 0.0
5:00 197.7 F 0.0
10:00 199.9 F 0.0
15:00 199.8 F 0.0
20:00 191.3 F 0.0
25:00 185.2 F 0.0
30:00 165.9 F 0.0
Drying Complete

FILLING: 1:14
HEATING: 24:32
STERILIZING: 4:00
VENTING: 3:43
DRYING: 30:00
TOTAL CYCLE: 01:03:28
CYCLE COMPLETE

10/04/2023 01:48 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____