

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2873

Name: RITTER2  
Sterilizer ID: V2313228  
09/12/2023 03:55 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.0 F 0.0  
2:00 173.3 F 3.2  
4:00 195.6 F 7.7  
6:00 211.4 F 12.7  
8:00 223.7 F 8.0  
10:00 235.1 F 13.2  
12:00 237.6 F 10.9  
14:00 236.7 F 9.6  
16:00 250.1 F 16.0  
18:00 261.3 F 22.5  
20:00 271.6 F 29.4

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 29.3  
1:00 272.2 F 30.0  
2:00 272.2 F 29.5  
3:00 271.3 F 29.5  
4:00 272.8 F 30.0  
  
Min 271.2 F 29.2  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 195.9 F 0.0  
5:00 202.4 F 0.0  
10:00 203.0 F 0.0  
15:00 205.2 F 0.0  
20:00 210.5 F 0.0  
25:00 212.6 F 0.0  
30:00 213.1 F 0.0  
Drying Complete

FILLING: 2:15  
HEATING: 20:30  
STERILIZING: 4:00  
VENTING: 3:40  
DRYING: 30:00  
TOTAL CYCLE: 01:00:26  
CYCLE COMPLETE

09/12/2023 04:55 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_