
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 000357
12:23 20/01/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:23
PRESSURIZING 2:34
STERILIZING 3:40
136.4 C 220kPa 3:40
136.4 C 233kPa 4:10
136.2 C 231kPa 4:40
136.4 C 233kPa 5:10
136.3 C 232kPa 5:40
136.1 C 231kPa 6:10
136.3 C 232kPa 6:40
136.4 C 233kPa 7:10
136.4 C 233kPa 7:10
VENTING 7:10
AIR DRYING 8:02
CYCLE COMPLETE 9:15

Digital Signature #
6716E0455854C76F
