
Midmark M11 - v1.0.0
Total Cycles: 5441

Name: RITTER 1
Sterilizer ID: V2178198
03/25/2025 10:25 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 181.6 F 0.0
2:00 174.2 F 2.5
4:00 175.1 F 6.4
6:00 201.5 F 10.2
8:00 213.9 F 4.1
10:00 225.7 F 7.9
12:00 234.8 F 12.5
14:00 235.0 F 9.2
16:00 240.5 F 10.3
18:00 245.1 F 13.1
20:00 256.6 F 19.2
22:00 266.9 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 28.5
1:00 271.9 F 29.2
2:00 272.4 F 29.2
3:00 273.1 F 29.5
4:00 272.2 F 29.1

Min 271.0 F 28.4
Max 273.3 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.0 F 0.0
5:00 220.7 F 0.0
10:00 221.5 F 0.0
15:00 210.1 F 0.0
20:00 210.0 F 0.0
25:00 211.6 F 0.0
30:00 212.9 F 0.0
Drying Complete

FILLING: 1:07
HEATING: 23:39
STERILIZING: 4:00
VENTING: 3:48
DRYING: 30:00
TOTAL CYCLE: 01:02:34
CYCLE COMPLETE

03/25/2025 11:27 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____