

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 004529  
13:15 14/10/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:37  
PRESSURIZING 2:40  
STERILIZING 3:52  
136.4 C 223kPa 3:52  
136.4 C 225kPa 4:23  
136.4 C 225kPa 4:53  
136.0 C 221kPa 5:23  
136.0 C 222kPa 5:53  
136.3 C 225kPa 6:23  
136.4 C 222kPa 6:53  
136.3 C 222kPa 7:23  
136.3 C 222kPa 7:23  
VENTING 7:23  
AIR DRYING 8:09  
CYCLE COMPLETE 14:44

-----  
Digital Signature #  
94AE1E3E0584BE77  
-----