

-----  
Midmark M11 - v1.0.0  
Total Cycles: 861

Name: RITTER 1  
Sterilizer ID: V2178198  
06/01/2021 02:57 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 180.7 F 0.0  
2:00 182.1 F 2.4  
4:00 183.4 F 5.9  
6:00 203.7 F 10.5  
8:00 222.1 F 5.3  
10:00 230.2 F 9.1  
12:00 238.8 F 10.8  
14:00 241.4 F 11.6  
16:00 239.5 F 10.0  
18:00 252.3 F 16.3  
20:00 263.3 F 22.7  
22:00 272.4 F 28.7

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.3  
1:00 272.4 F 28.9  
2:00 271.3 F 28.2  
3:00 272.2 F 29.0  
4:00 272.0 F 28.6

Min 271.1 F 28.1  
Max 272.5 F 29.0

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 208.4 F 0.0  
5:00 234.3 F 0.0  
10:00 224.0 F 0.0  
15:00 199.9 F 0.0  
20:00 193.2 F 0.0  
25:00 215.0 F 0.0  
30:00 231.3 F 0.0  
Drying Complete

FILLING: 1:32  
HEATING: 22:17  
STERILIZING: 4:00  
VENTING: 3:26  
DRYING: 30:00  
TOTAL CYCLE: 01:01:15  
CYCLE COMPLETE

06/01/2021 03:58 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_