
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 004798
11:24 05/12/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:55
PRESSURIZING 2:58
STERILIZING 4:00
136.5 C 222kPa 4:00
136.4 C 224kPa 4:31
136.4 C 222kPa 5:01
136.4 C 224kPa 5:31
136.4 C 223kPa 6:01
136.4 C 225kPa 6:31
136.4 C 224kPa 7:01
136.1 C 221kPa 7:31
136.1 C 221kPa 7:31
VENTING 7:31
AIR DRYING 8:22
CYCLE COMPLETE 16:49

Digital Signature #
540A3137E85A4302
