
Midmark M11 - v1.0.0
Total Cycles: 2387

Name: RITTER 1
Sterilizer ID: V2178198
10/03/2022 11:36 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 200.3 F 0.0
2:00 197.9 F 3.7
4:00 206.9 F 8.7
6:00 224.3 F 8.7
8:00 232.7 F 10.3
10:00 232.6 F 7.5
12:00 243.4 F 11.9
14:00 245.8 F 13.1
16:00 259.2 F 20.3
18:00 270.9 F 27.9

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.4
1:00 272.0 F 28.5
2:00 272.6 F 29.1
3:00 271.5 F 28.4
4:00 272.2 F 28.7

Min 271.5 F 28.2
Max 272.9 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.1 F 0.0
5:00 223.2 F 0.0
10:00 223.2 F 0.0
15:00 235.3 F 0.0
20:00 226.9 F 0.0
25:00 224.2 F 0.0
30:00 222.3 F 0.0
Drying Complete

FILLING: 2:12
HEATING: 18:40
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:01
CYCLE COMPLETE

10/03/2022 12:33 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____