
Midmark M11 - v1.0.0
Total Cycles: 3511

Name: RITTER 1
Sterilizer ID: V2178198
07/11/2023 10:20 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.2 F 0.0
2:00 187.2 F 2.5
4:00 187.0 F 6.2
6:00 209.6 F 10.8
8:00 220.7 F 5.3
10:00 231.3 F 10.7
12:00 232.1 F 7.4
14:00 244.5 F 13.1
16:00 243.3 F 12.2
18:00 255.8 F 18.6
20:00 266.6 F 25.2

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.5
1:00 272.5 F 29.1
2:00 271.4 F 28.4
3:00 272.5 F 29.3
4:00 271.8 F 28.6

Min 271.3 F 28.3
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 205.9 F 0.0
5:00 209.7 F 0.0
10:00 225.3 F 0.0
15:00 229.2 F 0.0
20:00 229.9 F 0.0
25:00 229.9 F 0.0
30:00 229.8 F 0.0
Drying Complete

FILLING: 1:28
HEATING: 21:35
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:00:19
CYCLE COMPLETE

07/11/2023 11:20 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____