
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.0uS / 0.0ppm
CYCLE NUMBER 005760
8:09 15/06/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:34
PRESSURIZING 2:37
STERILIZING 3:30
136.0 C 220kPa 3:30
136.3 C 223kPa 4:00
136.4 C 223kPa 4:30
136.4 C 222kPa 5:00
136.0 C 221kPa 5:30
136.2 C 220kPa 6:00
136.3 C 223kPa 6:30
136.1 C 222kPa 7:00
136.1 C 222kPa 7:00
VENTING 7:00
AIR DRYING 7:50
CYCLE COMPLETE 8:02

Digital Signature #
CAB0FFE1CEC29819
