
Midmark M11 - v1.0.0
Total Cycles: 634

Name: MIDMARK
Sterilizer ID: V2178198
01/27/2021 09:55 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 113.7 F 0.0
2:00 111.6 F 0.0
4:00 127.9 F 1.4
6:00 155.9 F 4.3
8:00 186.0 F 8.6
10:00 229.5 F 8.5
12:00 221.7 F 7.7
14:00 235.2 F 12.8
16:00 236.5 F 10.0
18:00 234.4 F 7.9
20:00 249.9 F 15.0
22:00 262.3 F 22.1
24:00 272.3 F 28.4

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 28.5
1:00 271.9 F 28.4
2:00 271.7 F 28.5
3:00 271.6 F 28.6
4:00 271.6 F 28.6

Min 271.5 F 28.4
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 231.5 F 0.0
10:00 226.6 F 0.0
15:00 212.7 F 0.0
20:00 235.5 F 0.0
25:00 231.9 F 0.0
30:00 225.4 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 24:38
STERILIZING: 4:00
VENTING: 2:56
DRYING: 30:00
TOTAL CYCLE: 01:02:34
CYCLE COMPLETE

01/27/2021 10:57 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____