
Midmark M11 - v1.0.0
Total Cycles: 2901

Name: RITTER 1
Sterilizer ID: V2178198
02/14/2023 03:53 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 195.9 F 0.0
2:00 194.4 F 4.7
4:00 198.2 F 9.8
6:00 217.8 F 6.9
8:00 230.1 F 9.9
10:00 232.1 F 8.0
12:00 243.3 F 13.0
14:00 242.5 F 11.7
16:00 255.8 F 18.5
18:00 267.1 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.5
1:00 272.2 F 29.1
2:00 272.2 F 28.8
3:00 271.5 F 28.5
4:00 272.7 F 29.3

Min 271.4 F 28.3
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.0 F 0.0
5:00 221.3 F 0.0
10:00 240.0 F 0.0
15:00 231.7 F 0.0
20:00 226.8 F 0.0
25:00 224.0 F 0.0
30:00 223.1 F 0.0
Drying Complete

FILLING: 1:53
HEATING: 19:20
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:58:29
CYCLE COMPLETE

02/14/2023 04:51 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____