
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000742
10:49 23/09/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:26
PRESSURIZING 2:33
STERILIZING 3:43
136.4 C 232kPa 3:43
136.2 C 231kPa 4:13
136.4 C 233kPa 4:43
136.2 C 231kPa 5:13
136.1 C 230kPa 5:43
136.4 C 232kPa 6:13
136.4 C 233kPa 6:43
136.4 C 233kPa 7:13
136.4 C 233kPa 7:13
VENTING 7:13
AIR DRYING 8:06
CYCLE COMPLETE 8:14

Digital Signature #
2317BEC236AA4FEA
