
Midmark M11 - v1.0.0
Total Cycles: 1381

Name: RITTER 1
Sterilizer ID: V2178198
11/18/2021 08:01 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.3 F 0.0
2:00 76.7 F 0.0
4:00 128.5 F 0.0
6:00 148.8 F 0.0
8:00 158.4 F 0.1
10:00 188.5 F 0.4
12:00 212.0 F 2.8
14:00 221.9 F 5.5
16:00 229.9 F 8.3
18:00 239.0 F 12.0
20:00 233.3 F 8.1
22:00 245.3 F 13.5
24:00 242.6 F 11.6
26:00 239.9 F 10.2
28:00 253.1 F 16.8
30:00 264.4 F 23.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 28.1
1:00 271.8 F 28.4
2:00 272.6 F 28.9
3:00 271.9 F 28.2
4:00 271.7 F 28.4

Min 271.4 F 27.9
Max 272.8 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.9 F 0.0
5:00 212.9 F 0.0
10:00 216.8 F 0.0
15:00 213.7 F 0.0
20:00 205.9 F 0.0
25:00 222.3 F 0.0
30:00 227.5 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 31:52
STERILIZING: 4:00
VENTING: 3:04
DRYING: 30:00
TOTAL CYCLE: 01:09:53
CYCLE COMPLETE

11/18/2021 09:10 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____