

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 005909  
15:20 05/07/2023  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:27  
PRESSURIZING 2:31  
STERILIZING 3:30  
136.5 C 223kPa 3:30  
136.1 C 223kPa 4:00  
136.4 C 225kPa 4:30  
136.2 C 223kPa 5:00  
136.2 C 223kPa 5:30  
136.4 C 222kPa 6:00  
136.2 C 224kPa 6:30  
136.2 C 221kPa 7:00  
136.2 C 221kPa 7:00  
VENTING 7:00  
AIR DRYING 7:52  
CASSETTE REMOVED  
CYCLE COMPLETE 24:04  
-----

Digital Signature #  
6F59E4E44EA297B6  
-----