
Midmark M11 - v1.0.5
Total Cycles: 425

Name: RITTER2
Sterilizer ID: V2313228
10/07/2021 09:41 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 83.1 F 0.0
2:00 83.8 F 0.0
4:00 90.6 F 0.2
6:00 115.2 F 2.4
8:00 150.0 F 5.6
10:00 167.7 F 9.5
12:00 190.0 F 11.0
14:00 210.1 F 5.9
16:00 222.7 F 9.8
18:00 230.9 F 10.7
20:00 236.5 F 10.8
22:00 234.3 F 8.6
24:00 247.7 F 14.8
26:00 258.4 F 20.7
28:00 268.2 F 27.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.8 F 29.9
2:00 271.5 F 29.3
3:00 272.8 F 29.8
4:00 271.5 F 29.3

Min 271.4 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 187.4 F 0.0
5:00 190.8 F 0.0
10:00 189.3 F 0.0
15:00 177.7 F 0.0
20:00 174.6 F 0.0
25:00 175.5 F 0.0
30:00 180.7 F 0.0
Drying Complete

FILLING: 0:60
HEATING: 29:18
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:07:25
CYCLE COMPLETE

10/07/2021 10:51 AM
ACCEPTED: X
REJECTED:
Operator: U02