
Midmark M11 - v1.0.5
Total Cycles: 3742

Name: RITTER2
Sterilizer ID: V2313228
06/27/2024 02:33 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 98.4 F 0.0
2:00 98.4 F 0.2
4:00 111.5 F 1.8
6:00 144.0 F 4.5
8:00 179.5 F 7.8
10:00 188.8 F 12.1
12:00 206.5 F 5.1
14:00 218.7 F 9.1
16:00 229.8 F 13.3
18:00 232.8 F 9.7
20:00 237.3 F 10.0
22:00 244.0 F 13.2
24:00 255.4 F 19.1
26:00 265.8 F 25.5

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 29.5
1:00 272.6 F 30.2
2:00 271.8 F 29.5
3:00 271.9 F 30.0
4:00 272.6 F 30.0

Min 271.3 F 29.3
Max 272.9 F 30.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.2 F 0.0
5:00 212.1 F 0.0
10:00 203.9 F 0.0
15:00 191.4 F 0.0
20:00 189.4 F 0.0
25:00 189.9 F 0.0
30:00 202.7 F 0.0
Drying Complete

FILLING: 0:57
HEATING: 27:50
STERILIZING: 4:00
VENTING: 3:36
DRYING: 30:00
TOTAL CYCLE: 01:06:23
CYCLE COMPLETE

06/27/2024 03:39 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____