

-----  
Midmark M11 - v1.0.5  
Total Cycles: 3745

Name: RITTER2  
Sterilizer ID: V2313228  
06/28/2024 11:37 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 131.3 F 0.0  
2:00 129.1 F 1.3  
4:00 141.2 F 4.1  
6:00 167.6 F 7.5  
8:00 193.8 F 11.3  
10:00 212.5 F 5.0  
12:00 217.6 F 9.7  
14:00 233.7 F 9.2  
16:00 237.3 F 11.0  
18:00 235.3 F 8.9  
20:00 248.3 F 15.2  
22:00 259.2 F 21.4  
24:00 269.7 F 28.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 29.2  
1:00 272.7 F 29.9  
2:00 271.2 F 29.1  
3:00 272.2 F 30.0  
4:00 272.2 F 29.6

Min 271.0 F 29.1  
Max 272.9 F 30.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 193.4 F 0.0  
5:00 220.9 F 0.0  
10:00 228.2 F 0.0  
15:00 223.6 F 0.0  
20:00 236.4 F 0.0  
25:00 229.0 F 0.0  
30:00 230.9 F 0.0  
Drying Complete

FILLING: 0:58  
HEATING: 25:00  
STERILIZING: 4:00  
VENTING: 3:38  
DRYING: 30:00  
TOTAL CYCLE: 01:03:36  
CYCLE COMPLETE

06/28/2024 12:40 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_