
Midmark M11 - v1.0.0
Total Cycles: 2232

Name: RITTER 1
Sterilizer ID: V2178198
08/16/2022 03:20 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 161.6 F 0.0
2:00 161.9 F 3.3
4:00 178.6 F 7.4
6:00 198.7 F 12.2
8:00 218.6 F 6.2
10:00 231.0 F 11.3
12:00 232.5 F 8.1
14:00 241.6 F 11.3
16:00 244.9 F 12.8
18:00 257.5 F 19.5
20:00 268.8 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.2 F 28.4
1:00 272.5 F 29.1
2:00 271.4 F 28.4
3:00 271.7 F 28.9
4:00 272.3 F 29.0

Min 271.1 F 28.2
Max 272.6 F 29.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 207.2 F 0.0
5:00 228.0 F 0.0
10:00 225.9 F 0.0
15:00 226.4 F 0.0
20:00 236.8 F 0.0
25:00 220.3 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 1:40
HEATING: 21:07
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 01:00:01
CYCLE COMPLETE

08/16/2022 04:19 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____