

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4530

Name: RITTER 1  
Sterilizer ID: V2178198  
05/07/2024 09:49 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 184.9 F 0.0  
2:00 181.3 F 2.3  
4:00 179.7 F 6.1  
6:00 193.9 F 10.0  
8:00 212.3 F 11.4  
10:00 218.3 F 7.2  
12:00 230.0 F 11.4  
14:00 231.6 F 7.3  
16:00 241.6 F 12.6  
18:00 240.1 F 10.7  
20:00 252.2 F 16.7  
22:00 262.9 F 22.9  
24:00 272.7 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.5  
1:00 272.5 F 29.2  
2:00 271.5 F 28.5  
3:00 272.0 F 29.1  
4:00 272.3 F 29.0

Min 271.2 F 28.4  
Max 272.5 F 29.3

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 210.5 F 0.0  
5:00 227.4 F 0.0  
10:00 228.4 F 0.0  
15:00 230.1 F 0.0  
20:00 230.7 F 0.0  
25:00 229.9 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 1:07  
HEATING: 24:22  
STERILIZING: 4:00  
VENTING: 3:48  
DRYING: 30:00  
TOTAL CYCLE: 01:03:17  
CYCLE COMPLETE

05/07/2024 10:52 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_