
Midmark M11 - v1.0.0
Total Cycles: 1024

Name: RITTER 1
Sterilizer ID: V2178198
07/15/2021 04:16 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.1 F 0.0
2:00 204.5 F 6.0
4:00 225.2 F 12.3
6:00 231.2 F 8.7
8:00 237.6 F 9.6
10:00 243.3 F 11.6
12:00 248.2 F 14.1
14:00 261.4 F 21.5
16:00 272.7 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.1
1:00 272.2 F 28.5
2:00 271.9 F 28.6
3:00 272.1 F 28.4
4:00 271.6 F 28.4

Min 271.3 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.7 F 0.0
5:00 222.8 F 0.0
10:00 190.5 F 0.0
15:00 183.1 F 0.0
20:00 177.6 F 0.0
25:00 173.6 F 0.0
30:00 170.6 F 0.0
Drying Complete

FILLING: 2:51
HEATING: 16:19
STERILIZING: 4:00
VENTING: 3:23
DRYING: 30:00
TOTAL CYCLE: 00:56:33
CYCLE COMPLETE

07/19/2021 06:22 AM
ACCEPTED: X
REJECTED:
Operator: U01