

-----  
Midmark M11 - v1.0.5  
Total Cycles: 2032

Name: RITTER2  
Sterilizer ID: V2313228  
02/09/2023 01:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 168.8 F 0.0  
2:00 169.9 F 1.9  
4:00 191.3 F 6.0  
6:00 210.9 F 11.0  
8:00 222.7 F 6.1  
10:00 236.0 F 11.7  
12:00 235.1 F 9.4  
14:00 236.5 F 9.1  
16:00 249.9 F 15.8  
18:00 262.5 F 22.9  
20:00 272.4 F 29.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 29.0  
1:00 272.2 F 29.5  
2:00 271.5 F 29.2  
3:00 272.3 F 29.6  
4:00 271.5 F 28.9

Min 271.3 F 28.9  
Max 272.4 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.2 F 0.0  
5:00 216.5 F 0.0  
10:00 231.4 F 0.0  
15:00 238.2 F 0.0  
20:00 230.6 F 0.0  
25:00 226.3 F 0.0  
30:00 218.8 F 0.0  
Drying Complete

FILLING: 1:05  
HEATING: 20:11  
STERILIZING: 4:00  
VENTING: 3:10  
DRYING: 30:00  
TOTAL CYCLE: 00:58:27  
CYCLE COMPLETE

02/09/2023 02:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_