
Midmark M11 - v1.0.5
Total Cycles: 228

Name: RITTER2
Sterilizer ID: V2313228
07/21/2021 06:22 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.6 F 0.0
2:00 81.3 F 0.1
4:00 75.5 F 0.7
6:00 98.1 F 2.7
8:00 129.3 F 5.1
10:00 156.7 F 7.9
12:00 180.9 F 11.4
14:00 209.7 F 4.0
16:00 209.3 F 8.3
18:00 222.6 F 12.5
20:00 227.9 F 8.9
22:00 238.1 F 11.6
24:00 241.2 F 12.1
26:00 252.6 F 18.0
28:00 263.4 F 24.2
30:00 272.7 F 30.3

STERILIZING
mm:ss Degrees PSI
0:00 272.4 F 30.4
1:00 272.1 F 30.7
2:00 272.3 F 30.6
3:00 272.1 F 30.5
4:00 272.3 F 30.4

Min 271.8 F 30.3
Max 272.9 F 30.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 178.9 F 0.0
5:00 228.5 F 0.0
10:00 222.7 F 0.0
15:00 227.8 F 0.0
20:00 226.2 F 0.0
25:00 213.6 F 0.0
30:00 219.7 F 0.0
Drying Complete

FILLING: 0:54
HEATING: 30:52
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:08:53
CYCLE COMPLETE

07/21/2021 07:31 AM
ACCEPTED: X
REJECTED:
Operator: U01