
Midmark M11 - v1.0.5
Total Cycles: 1302

Name: RITTER2
Sterilizer ID: V2313228
07/15/2022 06:45 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 74.1 F 0.0
2:00 76.3 F 0.0
4:00 107.8 F 0.0
6:00 125.4 F 0.3
8:00 150.8 F 0.6
10:00 179.5 F 3.9
12:00 189.1 F 7.7
14:00 204.1 F 11.4
16:00 217.1 F 5.0
18:00 228.1 F 9.4
20:00 236.3 F 11.7
22:00 237.9 F 10.9
24:00 235.4 F 8.8
26:00 248.4 F 15.0
28:00 259.1 F 21.0
30:00 269.0 F 27.4

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.2
1:00 272.4 F 29.7
2:00 271.0 F 29.2
3:00 272.3 F 29.8
4:00 271.1 F 29.0

Min 271.0 F 29.0
Max 272.5 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 185.5 F 0.0
5:00 196.0 F 0.0
10:00 202.9 F 0.0
15:00 202.9 F 0.0
20:00 206.9 F 0.0
25:00 213.8 F 0.0
30:00 211.8 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 31:08
STERILIZING: 4:00
VENTING: 3:07
DRYING: 30:00
TOTAL CYCLE: 01:09:11
CYCLE COMPLETE

07/15/2022 07:54 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____