
Midmark M11 - v1.0.5
Total Cycles: 956

Name: RITTER2
Sterilizer ID: V2313228
04/05/2022 09:48 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 193.3 F 0.0
2:00 191.9 F 0.5
4:00 201.3 F 4.8
6:00 216.8 F 10.5
8:00 225.9 F 7.0
10:00 241.1 F 13.2
12:00 241.3 F 11.9
14:00 241.5 F 11.3
16:00 255.2 F 18.5
18:00 267.2 F 25.8

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.9
1:00 272.7 F 29.6
2:00 271.9 F 28.8
3:00 272.3 F 29.6
4:00 272.6 F 29.3

Min 271.5 F 28.7
Max 272.9 F 29.7

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.3 F 0.0
5:00 232.5 F 0.0
10:00 224.6 F 0.0
15:00 228.8 F 0.0
20:00 229.0 F 0.0
25:00 230.7 F 0.0
30:00 224.9 F 0.0
Drying Complete

FILLING: 1:16
HEATING: 19:26
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:57:50
CYCLE COMPLETE

04/05/2022 10:45 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____