
Midmark M11 - v1.0.0
Total Cycles: 1831

Name: RITTER 1
Sterilizer ID: V2178198
04/18/2022 08:34 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 171.0 F 0.0
2:00 163.0 F 0.5
4:00 164.3 F 4.0
6:00 186.0 F 8.4
8:00 209.3 F 13.5
10:00 226.4 F 7.3
12:00 237.8 F 12.2
14:00 235.6 F 8.8
16:00 241.1 F 10.3
18:00 246.2 F 13.0
20:00 258.3 F 19.6
22:00 269.3 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.1 F 28.0
1:00 272.5 F 28.9
2:00 272.1 F 28.3
3:00 272.7 F 29.0
4:00 272.2 F 28.4

Min 271.0 F 27.8
Max 272.9 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.5 F 0.0
5:00 195.9 F 0.0
10:00 203.6 F 0.0
15:00 208.8 F 0.0
20:00 215.2 F 0.0
25:00 245.1 F 0.0
30:00 232.9 F 0.0
Drying Complete

FILLING: 0:50
HEATING: 23:09
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:01:04
CYCLE COMPLETE

04/18/2022 09:35 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____