

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2804

Name: RITTER 1  
Sterilizer ID: V2178198  
01/23/2023 11:42 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 202.6 F 0.0  
2:00 188.7 F 1.5  
4:00 193.0 F 5.7  
6:00 208.4 F 10.1  
8:00 232.3 F 7.4  
10:00 228.9 F 10.5  
12:00 234.7 F 7.8  
14:00 242.0 F 12.4  
16:00 244.1 F 12.7  
18:00 256.2 F 19.1  
20:00 266.9 F 25.9  
22:00 272.5 F 29.5

STERILIZING  
mm:ss Degrees PSI  
0:00 272.0 F 29.3  
1:00 272.7 F 29.6  
2:00 272.4 F 29.8  
3:00 271.7 F 29.5  
4:00 272.0 F 29.4

Min 271.6 F 29.2  
Max 272.7 F 29.8

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 194.4 F 0.0  
5:00 222.9 F 0.0  
10:00 238.9 F 0.0  
15:00 234.2 F 0.0  
20:00 228.1 F 0.0  
25:00 228.9 F 0.0  
30:00 230.6 F 0.0  
Drying Complete

FILLING: 1:28  
HEATING: 22:06  
STERILIZING: 4:00  
VENTING: 3:36  
DRYING: 30:00  
TOTAL CYCLE: 01:01:10  
CYCLE COMPLETE

01/23/2023 12:43 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_